

SOLUTION CHANGE

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 02/01/2024

OPERATOR NAME:
 RYAN S PEDONE
 Permit #: 35878
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:24
INTERNAL STANDARD	VERIFIED	09:24
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	09:25
BLANK TEST	0.00	09:26
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	09:26
BLANK TEST	0.00	09:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:27
BLANK TEST	0.00	09:28

