

SUPERVISOR TEST

GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106

Date: 02/19/2024

OPERATOR NAME:  
MICHAELA R ORTALE  
Permit #: 52289  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 23430

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 06:12 |
| INTERNAL STANDARD       | VERIFIED | 06:13 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.09     | 06:13 |
| BLANK TEST              | 0.00     | 06:14 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.09     | 06:14 |
| BLANK TEST              | 0.00     | 06:15 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.09     | 06:15 |
| BLANK TEST              | 0.00     | 06:16 |

