

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 02/05/2024

OPERATOR NAME:  
NATHANIEL L WARRINGTON  
Permit #: 52169  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:47
INTERNAL STANDARD	VERIFIED	01:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:48
BLANK TEST	0.00	01:49
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	01:49
BLANK TEST	0.00	01:50
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	01:50
BLANK TEST	0.00	01:51

