

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 03/04/2024

OPERATOR NAME:  
NATHANIEL L WARRINGTON  
Permit #: 52169  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 23430

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 03:51 |
| INTERNAL STANDARD       | VERIFIED | 03:51 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.10     | 03:51 |
| BLANK TEST              | 0.00     | 03:52 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.10     | 03:52 |
| BLANK TEST              | 0.00     | 03:53 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.10     | 03:54 |
| BLANK TEST              | 0.00     | 03:54 |

