

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 03/11/2024

OPERATOR NAME:  
HUNTER L BARTON  
Permit #: 49353  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:07
INTERNAL STANDARD	VERIFIED	05:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:07
BLANK TEST	0.00	05:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:09
BLANK TEST	0.00	05:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:10
BLANK TEST	0.00	05:11

