

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 03/18/2024

OPERATOR NAME:
HUNTER L BARTON
Permit #: 49353
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:56
INTERNAL STANDARD	0.10	05:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:56
BLANK TEST	0.00	05:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:57
BLANK TEST	0.00	05:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:58
BLANK TEST	0.00	05:59

