

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 03/25/2024

OPERATOR NAME:  
ALEXANDER F DAAB  
Permit #: 52281  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:14
INTERNAL STANDARD	VERIFIED	00:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:15
BLANK TEST	0.00	00:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:16
BLANK TEST	0.00	00:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:17
BLANK TEST	0.00	00:18

