

SOLUTION CHANGE

-----  
 GLENS FALLS CITY  
 POLICE DEPT  
 DATAMASTER dmt: 117106  
 -----

Date: 04/01/2024

OPERATOR NAME:  
 NICHOLAS J LEHOISKY  
 Permit #: 49359  
 Operator Agency/Dept: GLENS FALLS CITY PD  
 Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:06
INTERNAL STANDARD	VERIFIED	05:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:07
BLANK TEST	0.00	05:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:08
BLANK TEST	0.00	05:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:09
BLANK TEST	0.00	05:10

