

SUPERVISOR TEST

GLEN'S FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 04/15/2024

OPERATOR NAME:
SAMANTHA M MCCONNELL
Permit #: 47102
Operator Agency/Dept: GLEN'S FALLS CITY PD
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:55
INTERNAL STANDARD	VERIFIED	05:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:55
BLANK TEST	0.00	05:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:56
BLANK TEST	0.00	05:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:58
BLANK TEST	0.00	05:58

