

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 04/22/2024

OPERATOR NAME:  
NICHOLAS J LEHOISKY  
Permit #: 49359  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:25
INTERNAL STANDARD	VERIFIED	03:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:25
BLANK TEST	0.00	03:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:27
BLANK TEST	0.00	03:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:28
BLANK TEST	0.00	03:29

