

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 04/29/2024

OPERATOR NAME:  
HUNTER L BARTON  
Permit #: 49353  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:00
INTERNAL STANDARD	VERIFIED	04:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	04:01
BLANK TEST	0.00	04:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	04:02
BLANK TEST	0.00	04:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	04:03
BLANK TEST	0.00	04:04

