

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 04/09/2024

OPERATOR NAME:
MICHAELA R ORTALE
Permit #: 52289
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:16
INTERNAL STANDARD	VERIFIED	02:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:17
BLANK TEST	0.00	02:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:18
BLANK TEST	0.00	02:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:19
BLANK TEST	0.00	02:20

