

SOLUTION CHANGE

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 05/01/2024

OPERATOR NAME:
MICHAELA R ORTALE
Permit #: 52289
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:05
INTERNAL STANDARD	VERIFIED	04:05
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	04:05
BLANK TEST	0.00	04:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	04:07
BLANK TEST	0.00	04:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	04:08
BLANK TEST	0.00	04:09

