

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 05/06/2024

OPERATOR NAME:
HUNTER L BARTON
Permit #: 49353
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:25
INTERNAL STANDARD	VERIFIED	05:25
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	05:25
BLANK TEST	0.00	05:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:26
BLANK TEST	0.00	05:27
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	05:28
BLANK TEST	0.00	05:28

