

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 05/12/2024

OPERATOR NAME:
HUNTER L BARTON
Permit #: 49353
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:02
INTERNAL STANDARD	VERIFIED	05:02
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	05:02
BLANK TEST	0.00	05:03
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	05:03
BLANK TEST	0.00	05:04
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	05:05
BLANK TEST	0.00	05:06

