

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 05/20/2024

OPERATOR NAME:  
HUNTER L BARTON  
Permit #: 49353  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:42
INTERNAL STANDARD	VERIFIED	04:43
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	04:43
BLANK TEST	0.00	04:44
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	04:44
BLANK TEST	0.00	04:45
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	04:45
BLANK TEST	0.00	04:46

