

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 05/27/2024

OPERATOR NAME:
ALEXANDER F DAAB
Permit #: 52281
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:18
INTERNAL STANDARD	VERIFIED	02:18
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:19
BLANK TEST	0.00	02:20
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	02:20
BLANK TEST	0.00	02:21
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	02:21
BLANK TEST	0.00	02:22

