

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 06/10/2024

OPERATOR NAME:  
ALEXANDER F DAAB  
Permit #: 52281  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:10
INTERNAL STANDARD	VERIFIED	00:10
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	00:11
BLANK TEST	0.00	00:12
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	00:12
BLANK TEST	0.00	00:13
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	00:13
BLANK TEST	0.00	00:14

