

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 06/17/2024

OPERATOR NAME:
MICHAELA R ORTALE
Permit #: 52289
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:24
INTERNAL STANDARD	VERIFIED	04:24
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	04:25
BLANK TEST	0.00	04:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:26
BLANK TEST	0.00	04:27
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	04:27
BLANK TEST	0.00	04:28

