

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 06/24/2024

OPERATOR NAME:
HUNTER L BARTON
Permit #: 49353
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:22
INTERNAL STANDARD	VERIFIED	06:23
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	06:23
BLANK TEST	0.00	06:24
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	06:24
BLANK TEST	0.00	06:25
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	06:25
BLANK TEST	0.00	06:26

