

SOLUTION CHANGE

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 07/01/2024

OPERATOR NAME:
SAMANTHA M MCCONNELL
Permit #: 47102
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:11
INTERNAL STANDARD	VERIFIED	06:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:11
BLANK TEST	0.00	06:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:13
BLANK TEST	0.00	06:14
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:14
BLANK TEST	0.00	06:15

