

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 07/08/2024

OPERATOR NAME:
HUNTER S MCKENNA
Permit #: 52925
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:14
INTERNAL STANDARD	VERIFIED	06:14
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	06:15
BLANK TEST	0.00	06:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:16
BLANK TEST	0.00	06:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:18
BLANK TEST	0.00	06:18

