

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 07/15/2024

OPERATOR NAME:  
ALEXANDER F DAAB  
Permit #: 52281  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:46
INTERNAL STANDARD	VERIFIED	06:46
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	06:47
BLANK TEST	0.00	06:47
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	06:48
BLANK TEST	0.00	06:49
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	06:50
BLANK TEST	0.00	06:51

