

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 07/22/2024

OPERATOR NAME:
HUNTER S MCKENNA
Permit #: 52925
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:10
INTERNAL STANDARD	VERIFIED	06:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:11
BLANK TEST	0.00	06:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:12
BLANK TEST	0.00	06:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:14
BLANK TEST	0.00	06:15

