

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 07/29/2024

OPERATOR NAME:
HUNTER S MCKENNA
Permit #: 52925
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:20
INTERNAL STANDARD	VERIFIED	05:20
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	05:20
BLANK TEST	0.00	05:21
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	05:22
BLANK TEST	0.00	05:23
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	05:23
BLANK TEST	0.00	05:24

