

SOLUTION CHANGE

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 08/02/2024

OPERATOR NAME:
MICHAELA R ORTALE
Permit #: 52289
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:58
INTERNAL STANDARD	VERIFIED	02:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:59
BLANK TEST	0.00	03:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:00
BLANK TEST	0.00	03:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:01
BLANK TEST	0.00	03:02

