

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 08/12/2024

OPERATOR NAME:
HUNTER S MCKENNA
Permit #: 52925
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:11
INTERNAL STANDARD	VERIFIED	05:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:12
BLANK TEST	0.00	05:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:14
BLANK TEST	0.00	05:14
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:15
BLANK TEST	0.00	05:16

