

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 08/19/2024

OPERATOR NAME:
ALEXANDER F DAAB
Permit #: 52281
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:32
INTERNAL STANDARD	VERIFIED	05:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:32
BLANK TEST	0.00	05:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:34
BLANK TEST	0.00	05:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:35
BLANK TEST	0.00	05:36

