

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 08/26/2024

OPERATOR NAME:  
SAMANTHA M MCCONNELL  
Permit #: 47102  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:10
INTERNAL STANDARD	VERIFIED	06:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:11
BLANK TEST	0.00	06:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:12
BLANK TEST	0.00	06:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:13
BLANK TEST	0.00	06:14

