

SOLUTION CHANGE

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 09/01/2024

OPERATOR NAME:
 ALEXANDER F DAAB
 Permit #: 52281
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:08
INTERNAL STANDARD	VERIFIED	04:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:08
BLANK TEST	0.00	04:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:10
BLANK TEST	0.00	04:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:11
BLANK TEST	0.00	04:12

