

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 09/02/2024

OPERATOR NAME:  
HUNTER L BARTON  
Permit #: 49353  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:27
INTERNAL STANDARD	VERIFIED	06:27
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	06:28
BLANK TEST	0.00	06:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:29
BLANK TEST	0.00	06:30
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	06:30
BLANK TEST	0.00	06:31

