

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 09/09/2024

OPERATOR NAME:  
ALEXANDER F DAAB  
Permit #: 52281  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:44
INTERNAL STANDARD	VERIFIED	02:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:44
BLANK TEST	0.00	02:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:46
BLANK TEST	0.00	02:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:48
BLANK TEST	0.00	02:48

