

SOLUTION CHANGE

-----  
 GLENS FALLS CITY  
 POLICE DEPT  
 DATAMASTER dmt: 117106  
 -----

Date: 10/24/2024

OPERATOR NAME:  
 RYAN S PEDONE  
 Permit #: 35878  
 Operator Agency/Dept: GLENS FALLS CITY PD  
 Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:04
INTERNAL STANDARD	VERIFIED	15:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	15:04
BLANK TEST	0.00	15:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	15:06
BLANK TEST	0.00	15:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	15:07
BLANK TEST	0.00	15:08

