

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 10/28/2024

OPERATOR NAME:  
ALEXANDER F DAAB  
Permit #: 52281  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:48
INTERNAL STANDARD	VERIFIED	01:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:48
BLANK TEST	0.00	01:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:50
BLANK TEST	0.00	01:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:51
BLANK TEST	0.00	01:52

