

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 11/04/2024

OPERATOR NAME:
HUNTER S MCKENNA
Permit #: 52925
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:15
INTERNAL STANDARD	VERIFIED	03:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:16
BLANK TEST	0.00	03:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:17
BLANK TEST	0.00	03:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:18
BLANK TEST	0.00	03:19

