

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 11/11/2024

OPERATOR NAME:  
NICHOLAS J LEHOISKY  
Permit #: 49359  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:49
INTERNAL STANDARD	VERIFIED	05:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:49
BLANK TEST	0.00	05:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:50
BLANK TEST	0.00	05:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:52
BLANK TEST	0.00	05:52

