

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 11/18/2024

OPERATOR NAME:  
ALEXANDER F DAAB  
Permit #: 52281  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:47
INTERNAL STANDARD	VERIFIED	00:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:48
BLANK TEST	0.00	00:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:49
BLANK TEST	0.00	00:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:50
BLANK TEST	0.00	00:51

