

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 11/25/2024

OPERATOR NAME:  
HUNTER S MCKENNA  
Permit #: 52925  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:53
INTERNAL STANDARD	VERIFIED	03:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:53
BLANK TEST	0.00	03:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:54
BLANK TEST	0.00	03:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:56
BLANK TEST	0.00	03:56

