

SOLUTION CHANGE

GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106

Date: 12/01/2024

OPERATOR NAME:  
MICHAELA R ORTALE  
Permit #: 52289  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:32
INTERNAL STANDARD	VERIFIED	03:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:32
BLANK TEST	0.00	03:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:34
BLANK TEST	0.00	03:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:35
BLANK TEST	0.00	03:36

