

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 12/02/2024

OPERATOR NAME:
ALEXANDER F DAAB
Permit #: 52281
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:40
INTERNAL STANDARD	VERIFIED	03:40
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:40
BLANK TEST	0.00	03:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:41
BLANK TEST	0.00	03:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:42
BLANK TEST	0.00	03:43

