

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 12/23/2024

OPERATOR NAME:
REESE P WHEATON
Permit #: 53611
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 53283

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:14
INTERNAL STANDARD	VERIFIED	03:14
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:14
BLANK TEST	0.00	03:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:16
BLANK TEST	0.00	03:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:17
BLANK TEST	0.00	03:18

