

SUPERVISOR TEST

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 12/30/2024

OPERATOR NAME:
 ALEXANDER F DAAB
 Permit #: 52281
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:07
INTERNAL STANDARD	VERIFIED	01:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:08
BLANK TEST	0.00	01:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:09
BLANK TEST	0.00	01:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:10
BLANK TEST	0.00	01:11

