

SOLUTION CHANGE

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 01/01/2025

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:12
INTERNAL STANDARD	VERIFIED	10:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:13
BLANK TEST	0.00	10:14
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:14
BLANK TEST	0.00	10:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:15
BLANK TEST	0.00	10:16

