

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 01/20/2025

OPERATOR NAME:
REESE P WHEATON
Permit #: 53611
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:58
INTERNAL STANDARD	VERIFIED	04:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:58
BLANK TEST	0.00	04:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:59
BLANK TEST	0.00	05:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:01
BLANK TEST	0.00	05:01

