

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 01/05/2025

OPERATOR NAME:
ANDREW J MARANVILLE
Permit #: 53581
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:24
INTERNAL STANDARD	VERIFIED	23:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:25
BLANK TEST	0.00	23:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:26
BLANK TEST	0.00	23:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:27
BLANK TEST	0.00	23:28

