

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 02/03/2025

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:34
INTERNAL STANDARD	VERIFIED	07:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:35
BLANK TEST	0.00	07:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:36
BLANK TEST	0.00	07:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:37
BLANK TEST	0.00	07:38

