SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 02/10/2025

OPERATOR NAME:
HUNTER L BARTON
Permit #: 49353
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:33
INTERNAL STANDARD	VERIFIED	03:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:34
BLANK TEST	0.00	03:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:35
BLANK TEST	0.00	03:36
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:36
BLANK TEST	0.00	03:37

7W▼■Ç▬►▬