

SOLUTION CHANGE

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 03/01/2025

OPERATOR NAME:  
ANDREW J MARANVILLE  
Permit #: 53581  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:30
INTERNAL STANDARD	VERIFIED	04:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:31
BLANK TEST	0.00	04:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:32
BLANK TEST	0.00	04:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:33
BLANK TEST	0.00	04:34

