

SUPERVISOR TEST

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 03/10/2025

OPERATOR NAME:
 REESE P WHEATON
 Permit #: 53611
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:38
INTERNAL STANDARD	VERIFIED	05:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:39
BLANK TEST	0.00	05:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:40
BLANK TEST	0.00	05:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:41
BLANK TEST	0.00	05:42

