

SUPERVISOR TEST

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 03/24/2025

OPERATOR NAME:
 ANDREW J MARANVILLE
 Permit #: 53581
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:04
INTERNAL STANDARD	VERIFIED	00:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:04
BLANK TEST	0.00	00:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:05
BLANK TEST	0.00	00:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:07
BLANK TEST	0.00	00:08

